



Get Clarity On Generics

Cost-Effective CT & MRI Contrast Agents

**FRESENIUS
KABI**

[WATCH VIDEO](#)

AJNR

Ex Vivo Study of the Physical Effect of Coils on Pressure and Flow Dynamics in Experimental Aneurysms

Hans G. Boecher-Schwarz, Kurt Ringel, Laszlo Kopacz, Axel Heimann and Oliver Kempfski

This information is current as of August 1, 2025.

AJNR Am J Neuroradiol 2000, 21 (8) 1532-1536
<http://www.ajnr.org/content/21/8/1532>

Ex Vivo Study of the Physical Effect of Coils on Pressure and Flow Dynamics in Experimental Aneurysms

Hans G. Boecher-Schwarz, Kurt Ringel, Laszlo Kopacz, Axel Heimann, and Oliver Kempfski

BACKGROUND AND PURPOSE: Recent experimental studies and a few case reports reveal that coiling may not lead to permanent occlusion of aneurysms by an organized thrombus. Therefore, biologic long-term prognosis seems to be doubtful, and the physical effect of coils may be important. The purpose of this study was to investigate the physical effect of coils on pressure and flow dynamics in aneurysms.

METHODS: Bifurcation aneurysms were created in eight rabbits, explanted after 3 weeks, and tested in a model with pulsatile perfusion with 0.9% saline and heparinized blood. Before and after densely packing with coils, systemic and intraaneurysmal pressure, aneurysmal pulsation, and impact measurements were recorded.

RESULTS: The peak and shape of the pressure waves in the aneurysm and in the delivery system were not significantly different before and after coiling. Under physiological intraaneurysmal pressure (while being perfused with saline), significant reduction ($P = .022$) of aneurysmal wall pulsation after coil embolization was noted. Overall, the aneurysmal impact on surrounding structures was statistically unchanged after coiling. However, in a few cases, after coil embolization, the observed increase of impact was more than doubled compared with the original values before coiling.

CONCLUSION: Coils do not physically affect intraaneurysmal pressure. After coiling, there is no significant reduction of flow rates through the aneurysm and no reduction of aneurysmal impact, but aneurysmal wall pulsation may be decreased.

Endovascular coil embolization seems to have become an established method in the treatment of saccular intracranial aneurysms. Although permanent occlusion rate, and therefore long-term prognosis, may be doubtful (1–4), the number of patients treated with coils increases each year. Radiologic and clinical reports of the results of coil embolization have been published during the last 6 years (5–10). Knowledge of the physical effects of coils on pressure and flow dynamics in saccular bifurcation aneurysms, however, is still lacking. Because intraaneurysmal pressure is considered to be an important factor in aneurysm growth and rupture (11), some experimental studies assess pres-

sure inside saccular aneurysms in relation to the parent artery. In well-created bifurcation aneurysms, no pressure difference was observed under normal conditions (12–14). Byrne and Guglielmi (15) assumed that coils absorb parts of the energy of the systolic pulse to the walls of the aneurysm, buffering the transmitted blood pressure before the systolic pulse wave hits the wall of the aneurysm. Novak et al (14) measured intraaneurysmal pressure before and after incomplete coiling (there was only one coil inside the sac) but did not observe any pressure reduction. They did observe a decrease in aneurysmal pulsation. The purpose of the present experimental study was to investigate the physical effect of densely packed coils on intraaneurysmal pressure and on aneurysmal pulsation. Moreover, the aneurysmal impact (the pulse-dependent development of force between the aneurysmal sac and the surrounding structures) was determined. To avoid proximal arterial stenoses (eg, spasms) and the effects of thrombosis, an ex vivo model was chosen.

Methods

Surgical Procedure

In eight male New Zealand albino rabbits (weight, 3–4 kg), bifurcation aneurysms were microsurgically created according

Received June 7, 1999; accepted after revision February 21, 2000.

From the Departments of Neurosurgery (H.G.B.-S.) and Neuroradiology (K.R.), University Hospital, and the Institut of Neurosurgical Pathophysiology (L.K., A.H., O.K.), University of Mainz, Mainz, Germany.

This work was presented in part at the ASNR/ASHNR/ASPNR/ASITN/ASSR Joint Meeting, San Diego, CA, May 1999.

This work was supported by a grant from the Friedhelm Frees Foundation, Wiesbaden, Germany.

Address reprint requests to Hans G Boecher-Schwarz, MD, Department of Neurosurgery, University Hospital, Langenbeckstr. 1, 55131 Mainz, Germany.




FIG 1. Schematic drawing of the experimental system. Systemic flow was obtained from a conventional peristaltic pump, and then four air chambers and one check valve were used to produce a nearly constant, laminar, and unidirectional flow. This flow was then superimposed with a pulsatile flow generated by an electronically controlled valve and a resistance system. Intraaneurysmal and systemic pressure were recorded online.

to the technique presented by Forrest and O'Reilly (16) and modified by Spetzger et al (1). The aneurysms were produced at a surgically created bifurcation of the common carotid artery by means of a venous graft pouch taken from the confluence of the right internal and external jugular vein. After 3 weeks, the aneurysms (final diameter, approximately 1 cm) were explanted.

Ex Vivo Model

Figure 1 shows a schematic drawing of the experimental system. An influx of 120 mL/min was obtained from a conventional peristaltic pump, and four air chambers and one check valve were used to produce a nearly constant, laminar, and unidirectional flow. This flow was then superimposed with a pulsatile flow generated by an electronically controlled valve (frequency generator PM 5139, Philips) and a resistance system. This led to a pump-independent aneurysmal systolic inflow and a passive diastolic outflow.

The pressure in the delivery system (systemic pressure) was recorded online by a transducer (BPT 5326; World Precision Instruments, Sarasota, FL). The intraaneurysmal pressure was recorded online by a Camino probe (Camino Laboratories, San Diego, CA). Aneurysmal pulsation was measured without contact by a laser displacement sensor (No. LAA-20-2-670-1000; Di-el, Umbach, Germany).

Impact measurements were obtained by using a mechanical strain gauge dynamometer (Statham force transducer, ± 30 g). We recorded the pulsatile force of the aneurysm sac on a round rigid rod. The rod had a diameter of 1 mm and was mounted at a right angle onto the surface of the aneurysm with a constant preload of 0.6 g. Before all experiments, the strain gauge dynamometer was calibrated to correlate rod displacement with force. The rod displacement necessary for the determination of force was small and well below wall displacement determined without contact by laser ($<10\%$). Therefore, our impact/force measurement can be regarded as being independent of wall displacement.

In vitro tests were performed by pulsatile perfusion with (first) 0.9% saline and (second) heparinized rabbit blood. Pulse frequency was always 180/min, which is close to the physiological frequency (130–200/min) of the rabbit heart (17). At the beginning of each experiment (during the first run with saline), the systolic and diastolic intraaneurysmal pressures were adjusted close to the physiological values (systolic, 85–140 mm Hg; diastolic, 50–80 mm Hg) of the rabbit circulation (18). The selected valve and resistance adjustments then remained unchanged during the entire experiment. Because of the higher viscosity of heparinized blood, the pressure values automatically increased and were therefore higher than those with saline.

Aneurysmal pulsation and impact were recorded in at least five different locations (the same locations before and after coiling) on the aneurysm surface for 10 s each (a total of >50 pulses each). After saline and heparinized blood measurement, the aneurysm sac was densely packed (as much as possible) with Guglielmi detachable coils under angiographic control during perfusion with 0.9% saline. In all cases, an obliteration of more than 90% was achieved. Pulsation, impact, and intraaneurysmal and systemic pressure were then recorded again, as described above. The rate of data collection for pressure, pulsation, and impact measurements was always 100/s. For statistic evaluation, paired *t* test or signed rank test was used.

Results

The calculated mean waves of intraaneurysmal and systemic pressure before and after coiling are shown in Figure 2A for saline and in Figure 2B for blood. The peak and shape of the pressure waves in the aneurysmal dome and in the delivery system were statistically not significantly different before and after coiling (signed rank test, $P > .05\%$).




FIG 2. Intraaneurysmal and systemic pressure measurements (\pm standard error of mean) obtained before and after coiling. The amount and shape of the pressure waves were not significantly different before and after coil embolization. (x axis, each time interval = 10^{-1} s)

A, Circulating fluid, 0.9% saline.

B, Circulating fluid, heparinized blood.




FIG 3. Aneurysmal pulsation measurements (\pm standard error of mean) obtained before and after coil embolization. (x axis, each time interval = 10^{-1} s)

A, After coiling and while circulating a 0.9% saline solution, a statistically significant ($P = .022$) reduction of wall displacement was observed

B, No statistically significant reduction was observed with blood.

Because of technical reasons, laser displacement measurements obtained without contact were possible in only five experiments. Wall displacement (s) during pulsation (approximately 0.1–0.25 mm with 0.9% saline and 0.05–0.15 mm with blood) depends not only on intraaneurysmal pressure but also on the physical dimensions of the aneurysm (14). Because all specimens had different dimensions, relative values were used to define pulsation (maximal aneurysmal wall displacement without coils, $s_{\max} = 100\%$). In Figure 3, the mean curves of aneurysmal pulsation are shown. Only for 0.9% saline (Fig 3A) did the statistical analysis (t test) yield a significant difference ($P = .022$) for wall displacement before and after coiling. Figure 4 shows the calculated wall displacement/pressure curves before and after coiling with 0.9% saline and blood, respectively. All curves describe hysteresis loops with pulsation as a function of pressure. Wall displacement follows variations in pressure with delay. The arrows in Figure 4A indicate the direction of the hysteresis loop. At the point of identical pressure, an aneurysm is smaller during the ascending versus the descending portion of the hysteresis curve. The area within each hysteresis loop was measured with a planimeter. Even though the statistical analysis yields a significant difference for wall displacement before and after coiling for 0.9% saline and even though the hysteresis loops in Figure 4A seem to be different, the differences in planimetric areas within the hysteresis loops did not reach statistical significance in either fluid used (t test, $P = .134$ for 0.9% saline). Even the quotient of area and maximal wall displacement did not reveal a significant difference ($P > .05$).

Aneurysmal impact was measured in seven specimens. Statistically, there was no significant difference (t test, $P > .05$) between the measured force before (100%) and after coiling and perfusion with 0.9% saline or blood, respectively. In two cases, however, a distinct increase of impact of more than

twice the original values was measured after coil embolization. Figure 5 shows the results of impact measurements after coiling of all experiments with the circulating 0.9% saline fluid.

Discussion

Some experimental studies deal with the pressure inside saccular aneurysms in relation to its parent artery. In well-created bifurcation aneurysms, no pressure difference was observed under normal conditions (12–14). In our models, systemic and intraaneurysmal pressure differed with a nearly constant variation. This variation can easily be explained by the difference in the technique used for systemic and intraaneurysmal pressure measurement. There are only a few experimental studies dealing with experimental aneurysms and their flow dynamics after coiling (14, 19, 20). Gobin et al (20) and Novak et al (14) measured intraaneurysmal pressure in relation to the pressure in the parent artery after partial coil occlusion. They did not find any change, whether coils were present within the aneurysm. Our study confirms this finding in that we also did not observe any difference in intraaneurysmal pressure before and after coiling, even after densely packing the aneurysmal sac.

Novak et al (14) measured aneurysmal pulsation and intraaneurysmal pressure in sidewall and bifurcation aneurysms. In contrast to the present study, their aneurysms were only partially embolized with one coil. Moreover, they did not use a displacement sensor without contact. Similar to our study, the study by Novak et al (14) revealed a hysteresis in the relationship of intraaneurysmal pressure and pulsation. Already after partial coil embolization, they observed a reduction of aneurysmal pulsation and hysteresis loop. They assumed complete removal of hysteresis after further embolization.




FIG 4. Wall displacement/pressure curves measured before and after coil embolization describe hysteresis loops. Wall displacement follows changes in pressure with delay (arrows).

A, Circulating fluid, 0.9% saline.

B, Circulating fluid, heparinized blood.




FIG 5. Results of impact/force measurements obtained for all experiments after coiling and while circulating a 0.9% saline solution. After embolization, in two cases, an increase of impact of more than twice the original values was measured. (x axis, each time interval = 10^{-1} s; y axis, impact/force [F]; diastolic value of impact = 0%; systolic peak value of impact before coiling = 100%)

The areas within the hysteresis loops are characterized by pressure, wall pulsation, and time lag between these components. In the series with 0.9% saline in the present study, significant reduction of aneurysmal pulsation was also observed, but the differences in hysteresis did not reach statistical significance ($P = .134$). The main reason may be the small number of experiments ($n = 5$). For statistical analysis, the area within the loop was measured. The area was distinctly decreased after coiling in three of five experiments. In only one aneurysm was the value somewhat increased; however, the t test did not reveal a statistically significant difference. Although more experiments may be necessary, it is important to note that the present study revealed that even in a completely coiled aneurysm, the hysteresis between pulsation and pressure does not disappear. This is contrary to the assumption made by Novak et al (14).

We were unable to confirm the data regarding reduction of aneurysmal pulsation when blood was used to perfuse the aneurysms. In our models, an-

eurysm packing with Guglielmi detachable coils and contactless laser pulsation measurements were performed first during perfusion with 0.9% saline and "physiological" intraaneurysmal pressure values. Under these conditions, significant reduction of aneurysmal pulsation was observed. This finding is probably correct and can be explained by the higher tension of the aneurysmal wall due to the memory of the coil loops. Under "unphysiologically" elevated intraaneurysmal pressure values (such as in our series with blood perfusion), the tension of the aneurysmal wall was primarily high and the coil effect, therefore, may have been too small.

No significant difference could be determined between the calculated mean aneurysmal impact/force values to the surrounding structures before and after coiling. Because of the memory of the coils, the aneurysmal wall is under higher tension and thus stiffer than before. Therefore, impact may increase. However, in a few experiments, impact was distinctly higher after coiling. This finding may be important in patients with visual field defects that are due to aneurysms compressing the optic chiasma.

Coils do not affect the peak and shape of the pressure wave inside the aneurysmal dome, but at physiological pressures, aneurysmal pulsation seems to be smaller after coiling. Perhaps the reduced wall movement is an important factor in the development of thrombosis within the first hours after embolization (21). However, if there are really large spaces within the aneurysmal sac months after treatment (1-4), the physical effect of the presence of coils in the aneurysm may be unimportant and the mere presence does not prevent rupture. In addition to the physical effects of coils, there is a biologic reaction between the coils and the aneurysmal wall. Histologic investigations of experimental bifurcation aneurysms in rabbits 3 months after coiling (unpublished results) show a distinct increase of aneurysmal wall thickness due to

chronic inflammation. Similar findings in aneurysms in humans have recently been described by Bavinzski et al (22). It may be that this biologic effect prevents aneurysmal rupture after partial recanalization.

Conclusion

Coils do not physically affect the peak and shape of the pressure wave at the aneurysmal wall. Under physiological intraaneurysmal pressure, aneurysmal pulsation seems to be reduced after coiling. The aneurysmal impact/force on the surrounding structures is not significantly different before and after coil embolization. However, in individual experiments, even a distinct increase was measured after coiling.

References

1. Spetzger U, Reul J, Weis J, Bertalanffy H, Thron A, Gilsbach JM. **Microsurgically produced bifurcation aneurysms in a rabbit model for endovascular coil embolization.** *J Neurosurg* 1996; 85:488–495
2. Reul J, Weis J, Spetzger U, Konert T, Fricke C, Thron A. **Long-term angiographic and histopathologic findings in experimental aneurysms of the carotid bifurcation embolized with platinum and tungsten coils.** *AJNR Am J Neuroradiol* 1997;18:35–42
3. Manabe H, Fujita S, Hatayama T, Ohkuma H, Suzuki S, Yagihashi S. **Embolization of ruptured cerebral aneurysms with interlocking detachable coils in acute stage.** *Intervent Neuroradiol* 1997;3:49–63
4. Manabe H, Fujita S, Hatayama T, Suzuki S, Yagihashi S. **Rerupture of coil-embolized aneurysm during long-term observation.** *J Neurosurg* 1998;88:1096–1098
5. Byrne JV, Adams CBT, Kerr RSC, Molyneux AJ. **Endovascular treatment of inoperable intracranial aneurysms with platinum coils.** *Br J Neurosurg* 1995;9:585–592
6. Richling B, Gruber A, Bavinzski G, Killer M. **GDC-system embolization for brain aneurysms: location and follow up.** *Acta Neurochir (Wien)* 1995;134:177–183
7. Moret J, Pierot L, Boulin A, Castaings L, Rey A. **Endovascular treatment of anterior communicating artery aneurysms using Guglielmi detachable coils.** *Neuroradiology* 1996;38:800–805
8. Malisch TW, Guglielmi G, Viñuela F, et al. **Intracranial aneurysms treated with the Guglielmi detachable coil: midterm clinical results in a consecutive series of 100 patients.** *J Neurosurg* 1997;87:176–183
9. Cognard C, Weill A, Castaings L, Rey A, Moret J. **Intracranial berry aneurysms: angiographic and clinical results after endovascular treatment.** *Radiology* 1998;206:499–510
10. Byrne JV, Sohn MJ, Molyneux AJ. **Five-year experience in using coil embolization for ruptured intracranial aneurysms: outcomes and incidence of late rebleeding.** *J Neurosurg* 1999;90: 656–663
11. Sekhar LN, Heros RC. **Origin, growth and rupture of saccular aneurysms: a review.** *Neurosurgery* 1981;8:248–260
12. Özdamar N, Celebi G. **Pressure distribution on the wall of experimental aneurysms.** *Acta Neurochir (Wien)* 1978;45:27–34
13. Sekhar LN, Scialabassi RJ, Sun M, Blue HB, Wasserman JF. **Intraaneurysmal pressure measurements in experimental saccular aneurysms in dogs.** *Stroke* 1988;19:352–356
14. Novak P, Glikstein R, Mohr G. **Pulsation-pressure relationship in experimental aneurysms: observation of aneurysmal hysteresis.** *Neurol Res* 1996;18:377–382
15. Byrne JV, Guglielmi G. **Endovascular Treatment of Intracranial Aneurysms: Endovascular Treatments.** Berlin: Springer-Verlag; 1998:104–132
16. Forrest M, O'Reilly GV. **Production of experimental aneurysms at a surgically created bifurcation.** *AJNR Am J Neuroradiol* 1989;10:400–402
17. Scheunert A, Trautmann A. **Lehrbuch der Veterinär-Physiologie.** vol 6. Berlin: Verlag Paul Parey; 1976:478
18. Scheunert A, Trautmann A. **Lehrbuch der Veterinär-Physiologie.** vol 6. Berlin: Verlag Paul Parey; 1976:563
19. Graves VB, Strother CM, Partington CR, Rappe A. **Flow dynamics of lateral carotid artery aneurysms and their effects on coil and balloons: an experimental study in dogs.** *AJNR Am J Neuroradiol* 1992;13:189–196
20. Gobin YP, Counord JL, Flaud P, Duffaux J. **In vitro study of haemodynamics in a giant saccular aneurysm model: influence of flow dynamics in the parent vessel and effects of coil embolization.** *Neuroradiology* 1994;36:530–536
21. Reul J, Spetzger U, Weis J, von Buelow S, Ince A, Thron A. **The nature of early intraluminal thrombosis in terminal aneurysms occluded with Guglielmi detachable coils.** *Intervent Neuroradiol* 1998;4:39–48
22. Bavinzski G, Talazoglu V, Killer M, et al. **Gross and microscopic histopathological findings in aneurysms of the human brain treated with Guglielmi detachable coils.** *J Neurosurg* 1999;91: 284–293