



Discover Generics

Cost-Effective CT & MRI Contrast Agents



FRESENIUS
KABI

WATCH VIDEO

AJNR

Advantages and Pitfalls in 3T MR Brain Imaging: A Pictorial Review

Bernd L. Schmitz, Andrik J. Aschoff, Martin H.K. Hoffmann
and Georg Grön

AJNR Am J Neuroradiol 2005, 26 (9) 2229-2237
<http://www.ajnr.org/content/26/9/2229>

This information is current as
of June 23, 2025.

Advantages and Pitfalls in 3T MR Brain Imaging: A Pictorial Review

Bernd L. Schmitz, Andrik J. Aschoff, Martin H.K. Hoffmann, and Georg Grön

From its very beginning, MR as an imaging method suffered from inherently low signal intensity. A typical way to compensate for low signal intensity is the repetition of measurements, which causes long imaging times. Besides that, a variety of different approaches are used to increase signal intensity. Gradient systems have been substantially improved for new sequences with higher intrinsic signal intensity, for example in steady-state sequences like trueFISP (1). Radio-frequency (RF) coil design has advanced to arrays of smaller coils (2), which also improve the signal intensity. Contrast media for special applications have also been developed (3–5) to increase MR signal intensity. Although these developments are independent of field strength, the advance of a new generation of MR scanners with higher B_0 fields opens another approach to substantially increase signal intensity (6, 7). These scanners have recently become more widely available in clinical routine diagnostics (6). Increased signal intensity, however, is not the only effect of higher field strength (7–11), and associated effects necessitating changes in acquisition strategies have lead to a debate about the usefulness of higher field strength in clinical settings (12–14). The focus of this review is to illustrate some of the effects and limitations of higher B_0 imaging providing imaging examples from neuroradiologic applications and practical considerations on how to overcome some of these limitations.

Theoretical Advantages and Disadvantages of MR Imaging at 3T

Signal Intensity-to-Noise Ratio

In theory, the intensity of MR signal is linearly correlated with the strength of the static magnetic field B_0 (7). When compared with common methods of signal intensity enhancement (eg, increasing the number of excitations [NEX]), higher B_0 strength appears to be of clear advantage. Although relative




FIG 1. Comparison of the theoretical signal intensity increase when increasing NEX (gray line) or field strength (B_0 , black line). To reach the same signal intensity one can get from the doubling field strength, the acquisition time is 4 times longer (circles). Numbers along the x axis refer to levels of NEX and B_0 , respectively.




FIG 2. Theoretic increase in relative signal intensity for NEX and field strengths B_0 (gray lines) as in Fig 1. Note the relative increase in SAR related to field strength (black line). Although signal intensity linearly increases with field strengths, SAR increases with the square of field strengths. Values on the y axis are arbitrary units solely to demonstrate the relationship between the increases of different parameters in one graph.

signal intensity increase is linearly correlated with B_0 , it follows the square root of NEX only—ie, to achieve doubling in signal intensity at constant B_0 necessitates a 4-fold increase in acquisition time (Fig 1).

Larmor Frequency, Wavelength, and Specific Absorption Ratio

There are some effects associated when moving to higher field strength that, at first glance, appear to be counterproductive. The main problem is the increase

Received February 19, 2005; accepted after revision April 18.

From the Departments of Radiology (B.L.S., A.J.A., M.H.K.H.) and Psychiatry (G.G.), University Hospitals at Ulm, Ulm, Germany.

This work was presented in abstract form at the 90th scientific assembly and annual meeting of the Radiological Society of North America, November 28–December 3, 2004, Chicago, IL.

Address correspondence to Bernd L. Schmitz, Department of Radiology, University Hospitals at Ulm, Ulm, Germany.





FIG 3. Chemical shift artifacts at different bandwidths at 1.5 (*upper row*: TR, 172 ms; TE, 9 ms; field of view [FOV], $220 \times 220 \text{ mm}^2$; matrix, 256×192) and 3T (*lower row*: TR, 206 ms; TE, 10 ms; section thickness, 5 mm; FOV, $220 \times 220 \text{ mm}^2$; matrix, 256×192). Bandwidth increases from left to right (60, 120, 240, 480 Hz/pixel), resulting in chemical shift of 7.4, 3.7, 1.9, and 0.9 pixels, respectively, at 3T, and half of these values at 1.5T. Note the double line of the occipital subcutaneous tissue at lower bandwidths (between *white arrows*). Of note also is the increasing noise at higher bandwidths.

FIG 4. T2 turbo spin-echo (*left*) and hyperecho (*right*) of the same section position and with all other parameters kept equal (TR, 3810 ms; TE, 78 ms; matrix, 832×416 ; FOV, $220 \times 220 \text{ mm}^2$; echo train length [ETL] 9; section thickness, 5 mm; bandwidth, 145 Hz/pixel; flip angle, 120°), demonstrating that there is no difference in image contrast while SAR was significantly reduced.



in excitation frequency ω according to the Larmor equation:

$$\omega = \gamma \cdot B$$

Where γ denotes the gyromagnetic ratio (42.58 MHz/T) and B the magnetic field strength. Therefore, the resonance frequency increases from approximately 63.9 MHz at 1.5T to 127.8 MHz at 3T. For calculating the wavelength λ in water (speed of light $\lambda_0 = 300.000.000 \text{ m/s}$ divided by resonance fre-

quency), one has to take the dielectric constant of water ($\epsilon = 81$) into account:

$$\lambda = \frac{\lambda_0}{\sqrt{\epsilon}}$$

It follows that wavelength λ in water is reduced from 52 cm at 1.5T to 26 cm at 3T (7). The shorter wavelengths are substantially closer to natural body diameters, which results in an increase of shielding




FIG 5. Gradient echo T1 (left: TR, 311 ms; TE, 2.5 ms; section thickness, 5 mm; matrix, 512×256 ; FOV, 220×220 mm²; bandwidth, 465 Hz/pixel; flip angle, 90°) compared with T1 spin-echo (right: TR, 700 ms; TE, 10 ms; section thickness, 5 mm; matrix, 256×192 ; FOV, 220×220 mm²; bandwidth, 200 Hz/pixel; flip angle, 90°) in the same subject at 3T, which indicates higher contrast in gradient echo than spin-echo sequences at 3T.




FIG 6. Left, T1 spin-echo image at 1.5T (TR, 600 ms; TE, 14 ms; bandwidth, 90 Hz/pixel; section thickness, 5 mm; matrix, 256×192 ; FOV, 220×220 mm²; flip angle, 90°). Right, T1 spin-echo at 3T (TR, 700 ms; TE, 10 ms; section thickness, 5 mm; 19 sections; bandwidth, 200 Hz/pixel; matrix, 256×192 ; FOV, 220×220 mm²; flip angle, 90°), which is indicative of the reduced gray-to-white matter contrast at higher fields.

effects and interferences (15) from superimposed RF waves with complex effects on RF homogeneity (16). The associated problems are most obvious in abdominal and pelvic imaging. Possible solutions like RF shimming and improved coil design (17) are being evaluated (18).

Energy deposition in terms of specific absorption ratio (SAR) relates to the square of excitation frequency, which means that problems with SAR limitations scale with the square of B_0 field strength (Fig. 2).

Relationship between Chemical Shift, Bandwidth, Signal Intensity, and SAR

Effects of chemical shift scale linearly with B_0 . Therefore, fat and water resonance frequencies differ by 220 Hz at 1.5T and 440 Hz at 3T (19). Moreover, chemical shift (CS) is inversely related to the receiver bandwidth (BW, in Hz/pixel):

$$CS \sim \frac{1}{BW}$$

From the relation above, it follows that compensation for chemical shift effects at 3T is possible by measuring with doubled bandwidth at 3T relative to 1.5T (Fig 3). Signal intensity, however, is inversely related to the square root of bandwidth (19):

$$Signal \sim \frac{1}{\sqrt{BW}}$$

Consequently, increasing bandwidth reduces part of the extra signal intensity provided from imaging at 3T. In other words, doubling bandwidth to keep chemical shift effects constant between 1.5T and 3T reduces the original signal intensity gain obtained from 3T when compared with 1.5T from 200% to 141%.

Unfortunately, a higher bandwidth aggravates the problem of SAR exposure, but, because numerous other factors—including type and number of RF pulses,

FIG 7. Turbo inversion recovery (TIR) sequence at 1.5T (*left*: TR, 8770 ms; TE, 92 ms; TI, 300; matrix, 512×256 ; section thickness, 3 mm; FOV, 220×220 mm²; bandwidth, 130 Hz/pixel) and 3T (*right*: TR, 8890 ms; TE, 95 ms; TI, 300; all other parameters equal), which demonstrates clear depiction of gray and white matter at both field strengths.





FIG 8. Same axial section position of a TIR (TR, 7670 ms; TE, 68 ms; TI, 300; ETL, 15; matrix, 448×224 ; section thickness, 4.5 mm; bandwidth, 130 Hz/pixel; flip angle, 150°) and T1 sequence (magnetization-preparation rapid gradient echo TR, 1880 ms; TE, 3.7 ms; matrix, 256×230 ; FOV, 256×256 mm²; section thickness, 5 mm, reconstructed from 1.0 mm primarily; bandwidth, 160 Hz/pixel; flip angle, 8°) after contrast agent in a patient with a contrast-enhancing lesion. Note the absent contrast enhancement in the TIR image (white arrows in the T1 image).



flip angle, number of sections, echo train length, sequence design, coil design, patient positioning, number of sections, and additional pulses for fat saturation—affect SAR as well, a general prediction is not adequate.

Altogether, compensating increased chemical shift at 3T by means of increasing bandwidth appears to be an unsatisfactory solution. Because chemical shift is not a severe problem in brain imaging, larger shifts between water and fat can probably be tolerated. For other body regions, however, compensating mechanisms such as fat suppression are more often necessary.

Practical Considerations Combined with Imaging at 3T

Solutions for SAR Problems at Higher Field Strengths

An effective approach to reduce SAR problems at higher fields is related to changes of the excitation or

refocusing flip angle. Because energy deposition is proportional to the square of flip angle (20), even small reductions of the flip angle lead to significant SAR decreases. Most technicians and radiologists, however, are unsure about how flip angle variations may affect image contrast and therefore refrain from changing this parameter. Moreover flip-angle reductions (21) may involve a reduction in signal intensity, limiting the gain from higher field.

New sequences using hyperechoes (22) make use of the effect that only the central k space is relevant for image contrast (23). Consequently, these sequences automatically narrow the excitation flip angles only for the outer k-space lines while exciting the central k space with higher flip angles to achieve contrast (Fig 4).

The use of gradient echo sequences is a valuable alternative (Fig 5) to the approach discussed above to significantly reduce SAR. These sequences, however,




FIG 9. Same section position with spin-echo T1-weighted sequences at 1.5T (upper row: TR, 700 ms; TE, 10 ms; section thickness, 5 mm; matrix, 256×192 ; FOV, 220×220 mm²; bandwidth, 200 Hz/pixel) and 3T (lower row: same imaging parameters). Flip angles decreasing from left to right 130°, 110°, 90°, 70°, 50°. The lowest flip angle shows the best gray-to-white matter contrast. The effect is stronger at higher field.




FIG 10. Patient with multiple cavernomas inducing large susceptibility artifacts in T2* imaging at 3T, which is indicative of the high sensitivity for susceptibility effects at 3T (TR, 800 ms; TE, 26 ms; flip angle, 20°; bandwidth, 80 Hz/pixel; section thickness, 5 mm; matrix, 320×320 ; FOV, 220×220 mm²).

are prone to susceptibility artifacts especially at higher B_0 and longer TE, limiting their use at 3T.




FIG 11. Empirical data from a blocked finger tapping fMRI study at 1.5T and 3T comparing BOLD signal intensity strength in terms of statistical effect sizes (beta values from a linear regression equation; units are arbitrary) at different TEs (54 volumes scanned, a stimulus onset asynchrony of 16 volumes, and an epoch length of 8 volumes). Beta values were computed from scaled epseries within the general linear model by using SPM 99. There is a more than linear increase in BOLD signal intensity with higher field strength.

Very promising approaches are parallel imaging techniques (24–29) that permit the carry over of signal intensity to speed. The underlying concept is the reduction of phase encoding steps necessary for image acquisition. By this, the number of excitation pulses is reduced, thereby decreasing SAR. SNR, however, also decreases by the very same mechanism. So, for example, to achieve the same SAR level at 3T one yields at 1.5T, a speed-up factor of 4 is needed. This speed-up factor reduces SNR by 50%. Consequently, for this example, one ends up with the same SAR, the same resolution and the same SNR one can get from 1.5T, but acquisition is 4 times faster. Un-

FIG 12. Comparison of susceptibility artifacts at the skull base for 1.5T and 3T at different TEs, keeping all other parameters equal (TR, 4500 ms; isotropic voxel size, 3.3 mm; 48 sections; bandwidth, 2170 Hz/pixel). Artifacts are larger at any TE for 3T and increase with rising TE for both field strengths. At low TE, however, skull base artifacts are tolerable at 3T.





FIG 13. Same section position in axial EPI images of the same subject (TR, 4090 ms; TE, 54 ms; isotropic voxel size, 3.3 mm; 36 sections at 3.0T). Decrease of susceptibility effects with increasing bandwidth (in Hz/pixel) due to shorter echo spacing and duration of EPI readout. At low bandwidths (752 Hz/pixel, in this example), distortion effects from the frontal sinus are clearly obvious; however, very high bandwidths (4882 Hz/pixel) do not yield a significant further reduction of distortions that are already achieved with bandwidths at 2520 Hz/pixel.

fortunately, parallel imaging does not yet primarily turn signal intensity increase into higher spatial resolution.

T1 Spin-Echo Contrast

Apparently, gray-to-white matter contrast is reduced in spin-echo T1 imaging at 3T (30) when compared with 1.5T (Fig 6). There are several factors contributing to this observation. T1 times of gray and

white matter lengthen and converge at higher fields (31–33). Moreover, shielding effects induced by eddy currents prevent central parts of the image from being properly excited (16), which results in reduced signal intensity of the basal ganglia region. In addition, magnetization transfer effects are enhanced at higher B_0 , thus reducing signal intensity and contrast. There are several ways to compensate for these effects. For example, inversion recovery sequences ap-




FIG 14. Comparison of EPI with different voxel size regarding susceptibility artifacts at 3T. (A) Thinner sections (matrix, 64×64 ; TR 4500 ms; TE, 38 ms; bandwidth, 2170 Hz/pixel; FOV, 190×190 mm²; 53 sections; section thickness, 2 and 5 mm) and (B) smaller in-plane voxel sizes (matrix, 128×128 ; all other parameters same as in A) result in fewer susceptibility artifacts, especially in frontobasal regions.

pear very well suited if one is only interested in increasing gray to white matter contrast (Fig 7). The inversion pulse, however, interferes with visualization of contrast enhancement following gadolinium administration. Enhancing lesions may not be visible (Fig 8), because in inversion recovery sequences, unlike conventional T1 spin-echo sequences, the tissue with the shortest T1 does not necessarily exhibit the brightest signal intensity, depending on T1. Therefore, inversion recovery sequences are not quite use-

ful for comparative pre- and postcontrast T1 spin-echo imaging, regardless of B_0 . A different approach to increase gray-to-white matter contrast during T1-weighted spin-echo imaging at both field strengths is to reduce the excitation flip angle (34). Although this reduces SNR slightly, the gain in gray-to-white matter contrast is obvious and more pronounced at 3T (Fig 9). The effect can be predicted from theoretical calculations (35) by using known T1 and T2 relaxation times of gray and white matter (33) but is empirically




Fig 15. Maximum intensity projection of a time-of-flight angiography (TR, 28 ms; TE, 4.92 ms; matrix, 704×576 ; FOV, 163×200 mm²; 92 sections; section thickness, 0.75 mm; bandwidth, 105 Hz/pixel; flip angle, 25°) at 3T showing the clear depiction of even very small vessels.

larger than the theoretical predictions at 3T, probably because of shielding and magnetization transfer effects (34).

Susceptibility Effects at Higher Field Strength

Susceptibility effects increase with higher B_0 . Although this can cause artifacts, it is also of clear advantage for susceptibility-related imaging such as T2* sequences for detection of hemosiderin (Fig 10) relevant for detection of microbleeding in vascular encephalopathy (36). Cerebral hemorrhage is reliably detected at 3T (37).

Functional MR Imaging at Higher Field Strength

Functional MR imaging (fMRI) clearly benefits from higher field strengths (38–41). Because of increased susceptibility effects at 3T BOLD (blood oxygen level-dependent) signals increase with higher B_0 (Fig 11). Artifacts at air-bone interfaces can, however, be substantially aggravated at higher field strength, and their dependence from echo times (TE) at 3T is more apparent than at lower magnetic field strengths.

To achieve the same BOLD signal intensity at 1.5T and 3T, one can use a markedly shorter TE at 3T. In

turn, shorter TE permits shorter acquisition times and gives the possibility to increase the number of repetitions. Consequently, it is possible to increase the temporal resolution for fMRI (higher sampling rate of the hemodynamic response function) or increase the statistical power for signal intensity analyses. It is important to bear in mind that shorter TE are very useful in solving the problem of distortions at the skull base, which increase with increasing B_0 because of an increase of susceptibility effects in T2*-weighted imaging (Fig 12).

Further handles for optimizing echoplanar sequences to control for distortion effects at 3T are bandwidth and echo spacing (Fig 13). Reducing section thickness or increasing in-plane resolution (eg, 128×128 matrices) are additional options of note (Fig. 14A, -B).

MR Angiography at Higher Field Strength

MR angiography (MRA) is one of the most significantly improved MR techniques at higher field strength (42–46). Regarding time of flight (TOF) imaging (Fig 15), longer T1 times at 3T have the effect that the signal intensity inside the vessels is preserved even with thicker sections and in smaller vessels. Moreover, SNR is significantly increased, making higher resolutions possible within reasonable acquisition times (47). This results in better diagnostic quality—for example, with respect to intracranial aneurysms (48, 49).

Conclusion

MR imaging at 3T offers new potential because of a substantial increase in signal intensity provided by the higher magnetic field; however, associated changes—including increased SAR exposure, changed T1 and T2 relaxation times, decreased T1 tissue contrast, and increased susceptibility effects—render high-field MR imaging more compelling. Routine neuroradiologic imaging is feasible and may benefit from higher magnetic field strength, but appears to be more complicated than at lower field strength, and requires extended knowledge of the physical and technical requirements of MR imaging.

References

- Schultz E, Felix R. [Rapid magnetic resonance tomography sequences: theoretical principles and clinical imaging characteristics]. *Digitale Bilddiag* 1988;8:50–58
- de Zwart JA, Ledden PJ, van Gelderen P, et al. Signal-to-noise ratio and parallel imaging performance of a 16-channel receive-only brain coil array at 3.0 tesla. *Magn Reson Med* 2004;51:22–26
- Albert MS, Cates GD, Driehuys B, et al. Biological magnetic resonance imaging using laser-polarized ¹²⁹Xe. *Nature* 1994;370:199–201
- Middleton H, Black RD, Saam B, et al. MR imaging with hyperpolarized ³He gas. *Magn Reson Med* 1995;33:271–275
- Schad LR, Bachert P, Bock M, et al. Hyperpolarized gases: a new type of MR contrast agents? *Acta Radiol Suppl* 1997;412:43–46
- Lin W, An H, Chen Y, et al. Practical consideration for 3T imaging. *Magn Reson Imaging Clin N Am* 2003;11:615–639
- Schick F. Whole-body MRI at high field: technical limits and clinical potential. *Eur Radiol* 2005;15:946–959

8. Norris DG. **High field human imaging.** *J Magn Reson Imaging* 2003;18:519–529
9. Frayne R, Goodyear BG, Dickhoff P, et al. **Magnetic resonance imaging at 3.0 tesla: challenges and advantages in clinical neurological imaging.** *Invest Radiol* 2003;38:385–402
10. Baudendistel KT, Heverhagen JT, Knopp MV. [Clinical MR at 3 tesla: current status.] *Radiologe* 2004;44:11–18
11. Kim DS, Garwood M. **High-field magnetic resonance techniques for brain research.** *Curr Opin Neurobiol* 2003;13:612–619
12. Ross JS. **The high-field-strength curmudgeon.** *AJNR Am J Neuroradiol* 2004;25:168–169
13. Shapiro MD, Magee T, Williams D, Ramnath R, Ross JS. **The time for 3T clinical imaging is now.** *AJNR Am J Neuroradiol* 2004;25:1628–1629; author reply 1629
14. Pattany PM. **3T MR imaging: the pros and cons.** *AJNR Am J Neuroradiol* 2004;25:1455–1456
15. Collins CM, Liu W, Schreiber W, et al. **Central brightening due to constructive interference with, without, and despite dielectric resonance.** *J Magn Reson Imaging* 2005;21:192–196
16. Hoult DI, Phil D. **Sensitivity and power deposition in a high-field imaging experiment.** *J Magn Reson Imaging* 2000;12:46–67
17. Adriany G, Van de Moortele PF, Wiesinger F, et al. **Transmit and receive transmission line arrays for 7 tesla parallel imaging.** *Magn Reson Med* 2005;53:434–445
18. Vaughan JT, Adriany G, Snyder CJ, et al. **Efficient high-frequency body coil for high-field MRI.** *Magn Reson Med* 2004;52:851–859
19. Hashemi RH, Bradley WG, Lisanti CJ. *MRI: the basics.* Philadelphia: Lippincott Williams & Wilkins;2004:141–142
20. Hashemi RH, Bradley WG, Lisanti CJ. *MRI: the basics.* Philadelphia: Lippincott Williams & Wilkins;2004:39
21. Graf H, Schick F, Claussen CD, Seemann MD. **MR visualization of the inner ear structures: comparison of 1.5 tesla and 3 tesla images.** *Rofo* 2004;176:17–20
22. Hennig J, Scheffler K. **Hyperechoes.** *Magn Reson Med* 2001;46:6–12
23. Paschal CB, Morris HD. **K-space in the clinic.** *J Magn Reson Imaging* 2004;19:145–159
24. Pruessmann KP. **Parallel imaging at high field strength: synergies and joint potential.** *Top Magn Reson Imaging* 2004;15:237–244
25. Bammer R, Schoenberg SO. **Current concepts and advances in clinical parallel magnetic resonance imaging.** *Top Magn Reson Imaging* 2004;15:129–158
26. Larkman DJ, Atkinson D, Hajnal JV. **Artifact reduction using parallel imaging methods.** *Top Magn Reson Imaging* 2004;15:267–275
27. Blaimer M, Breuer F, Mueller M, et al. **SMASH, SENSE, PILS, GRAPPA: how to choose the optimal method.** *Top Magn Reson Imaging* 2004;15:223–236
28. Heidemann RM, Ozsarlak O, Parizel PM, et al. **A brief review of parallel magnetic resonance imaging.** *Eur Radiol* 2003; 13:2323–2337
29. Heidemann RM, Griswold MA, Muller M, et al. [Feasibilities and limitations of high field parallel MRI]. *Radiologe* 2004;44:49–55
30. Hu X, Norris DG. **Advances in high-field magnetic resonance imaging.** *Annu Rev Biomed Eng* 2004;6:157–184
31. Ethofer T, Mader I, Seeger U, et al. **Comparison of longitudinal metabolite relaxation times in different regions of the human brain at 1.5 and 3 tesla.** *Magn Reson Med* 2003;50:1296–1301
32. Michaeli S, Garwood M, Zhu XH, et al. **Proton T2 relaxation study of water, N-acetylaspartate, and creatine in human brain using Hahn and Carr-Purcell spin echoes at 4T and 7T.** *Magn Reson Med* 2002;47:629–633
33. Wansapura JP, Holland SK, Dunn RS, Ball WS Jr. **NMR relaxation times in the human brain at 3.0 tesla.** *J Magn Reson Imaging* 1999;9:531–538
34. Schmitz BL, Grön G, Brausewetter F, et al. **Enhancing gray-to-white matter contrast in 3T T1 spin-echo brain scans by optimizing flip angle.** *AJNR Am J Neuroradiol* 2005;26:2000–2004
35. Mills TC, Ortendahl DA, Hylton NM, et al. **Partial flip angle MR imaging.** *Radiology* 1987;162:531–539
36. Kato H, Izumiyama M, Izumiyama K, et al. **Silent cerebral microbleeds on T2*-weighted MRI: correlation with stroke subtype, stroke recurrence, and leukoaraiosis.** *Stroke* 2002;33:1536–1540
37. Allkemper T, Tombach B, Schwindt W, et al. **Acute and subacute intracerebral hemorrhages: comparison of MR imaging at 1.5 and 3.0 T: initial experience.** *Radiology* 2004;232:874–881
38. Hennig J, Speck O, Koch MA, Weiller C. **Functional magnetic resonance imaging: a review of methodological aspects and clinical applications.** *J Magn Reson Imaging* 2003;18:1–15
39. Chen W, Ugurbil K. **High spatial resolution functional magnetic resonance imaging at very-high-magnetic field.** *Top Magn Reson Imaging* 1999;10:63–78
40. Silvennoinen MJ, Clingman CS, Golay X, et al. **Comparison of the dependence of blood R2 and R2* on oxygen saturation at 1.5 and 4.7 tesla.** *Magn Reson Med* 2003;49:47–60
41. Yacoub E, Shmuel A, Pfeuffer J, et al. **Imaging brain function in humans at 7 tesla.** *Magn Reson Med* 2001;45:588–594
42. Bernstein MA, Huston J 3rd, Lin C, et al. **High-resolution intracranial and cervical MRA at 3.0T: technical considerations and initial experience.** *Magn Reson Med* 2001;46:955–962
43. Gaa J, Weidauer S, Requardt M, et al. **Comparison of intracranial 3D-ToF-MRA with and without parallel acquisition techniques at 1.5T and 3.0T: preliminary results.** *Acta Radiol* 2004;45:327–332
44. Willinek WA, Gieseke J, von Falkenhausen M, et al. **Sensitivity encoding (SENSE) for high spatial resolution time-of-flight MR angiography of the intracranial arteries at 3.0 T.** *Rofo* 2004;176:21–26
45. Willinek WA, Born M, Simon B, et al. **Time-of-flight MR angiography: comparison of 3.0-T imaging and 1.5-T imaging—initial experience.** *Radiology* 2003;229:913–920
46. Campeau NG, Huston J 3rd, Bernstein MA, et al. **Magnetic resonance angiography at 3.0 tesla: initial clinical experience.** *Top Magn Reson Imaging* 2001;12:183–204
47. Al-Kwif O, Emery DJ, Wilman AH. **Vessel contrast at three tesla in time-of-flight magnetic resonance angiography of the intracranial and carotid arteries.** *Magn Reson Imaging* 2002;20:181–187
48. Gibbs GF, Huston J 3rd, Bernstein MA, et al. **Improved image quality of intracranial aneurysms: 3.0-T versus 1.5-T time-of-flight MR angiography.** *AJNR Am J Neuroradiol* 2004;25:84–87
49. Gibbs GF, Huston J 3rd, Bernstein MA, et al. **3.0-tesla MR angiography of intracranial aneurysms: comparison of time-of-flight and contrast-enhanced techniques.** *J Magn Reson Imaging* 2005;21:97–102