



The Varying Porosity of Braided Self-Expanding Stents and Flow Diverters: An Experimental Study

A. Makoyeva, F. Bing, T.E. Darsaut, I. Salazkin and J. Raymond

This information is current as of June 21, 2025.

AJNR Am J Neuroradiol 2013, 34 (3) 596-602

doi: <https://doi.org/10.3174/ajnr.A3234>

<http://www.ajnr.org/content/34/3/596>

The Varying Porosity of Braided Self-Expanding Stents and Flow Diverters: An Experimental Study

A. Makoyeva, F. Bing, T.E. Darsaut, I. Salazkin, and J. Raymond

ABSTRACT

BACKGROUND AND PURPOSE: Braided self-expandable stents and flow diverters of uniform construction may develop zones of heterogeneous porosity in vivo. Unwanted stenoses may also occur at the extremities of the device. We studied these phenomena in dedicated benchtop experiments.

MATERIALS AND METHODS: Five braided devices of decreasing porosity were studied. To simulate discrepancies in diameters between the landing zones of the parent vessel and the aneurysm neck area, device extremities were inserted into silicone tubes of various diameters (2–3 mm), leaving the midportion free to react to experimental manipulations, which included axial approximation of the tubes (0–7 mm), and curvature (0–135°), with or without axial compression (0–2 mm). The length of the landing zone was sequentially decreased to study terminal device stenosis.

RESULTS: All devices adopted a conformation characterized by 3 different zones: bilateral landing zones, a middle compaction zone, and 2 transition zones. It is possible, during deployment, to compact stents and FDs to decrease porosity, but a limiting factor was the transition zone, which remained relatively unchanged and of higher porosity than the expansion zone. Length of the transition zone increased when devices were constrained in smaller tubes. Heterogeneities in porosity with compaction and curvatures were predictable and followed simple geometric rules. Extremity stenoses occurred increasingly with decreasing length of the landing zone.

CONCLUSIONS: Braided self-expandable devices show predictable changes in porosity according to device size, vessel diameter, and curvature. Adequate landing zones are required to prevent terminal device stenosis.

ABBREVIATIONS: FD = flow diverter; FSS = free segment of stent; HPS = high-porosity stent; P = porosity; SF = strut frequency

Some HPSs and most FDs used to treat intracranial aneurysms are constructed with multiple braided metal filaments to allow compression and loading into a microcatheter lumen. Once deployed into vascular structures of a larger diameter, these expand and shorten as they return toward their original size and conformation. We have previously shown in animal models that as these devices adapt to the constraints of local anatomy, the ensuing deformations can lead to substantial variations in device porosity between cases treated with the same device as well as wide varia-

tions in porosity along the different segments of the same device.^{1,2} Ideally, the aim of a flow-diversion treatment strategy is to form a tubular conduit of optimal porosity that will normalize linear flow, reconstruct the parent vessel, and effectively occlude the aneurysm while preserving arterial branches and perforators. Because the amount of “metallic coverage” of aneurysm and branch ostia may affect treatment safety and efficacy, it is desirable to accurately predict when and to what extent these deformations will occur and how these will influence device porosity.

We have previously tested both high-porosity braided self-expanding stents and flow diverters in wide-neck, bifurcation, and giant aneurysm models, and failures have occurred, particularly when devices were deployed in a curved configuration.^{1–3} Device deformations and subsequent heterogeneity in device porosity were frequent and sometimes quite substantial.^{1–3} When in vivo porosities of the portion of the devices bridging the aneurysm ostium were measured, it became clear that these were strongly influenced by the discrepancy in diameter between the parent artery and the portion of the device that was free to expand at the

Received March 5, 2012; accepted after revision May 16, 2012.

From the Department of Radiology (A.M., T.E.D., I.S., J.R.), Centre Hospitalier de l'Université de Montréal, Notre-Dame Hospital, Montreal, Quebec, Canada; Service de Radiologie A (F.B.), Centre Hospitalier de l'Université Strasbourg, Strasbourg 1, Hôpital Civil, Strasbourg, France; and Division of Neurosurgery, Department of Surgery (T.E.D.), University of Alberta Hospital, Mackenzie Health Sciences Centre, Edmonton, Alberta, Canada.

Please address correspondence to Jean Raymond, MD, CHUM–Notre-Dame Hospital, Interventional Neuroradiology (NRI), 1560 Sherbrooke East, Pavillon Simard, Room Z12909, Montreal, Quebec, Canada H2L 4M1; e-mail: jean.raymond@umontreal.ca

<http://dx.doi.org/10.3174/ajnr.A3234>

Device characteristics

	HPS	FD ₄₈	SSFD ₃₆	SSFD ₄₈	SSFD ₆₄
Nominal stent diameter × length (mm)	4.5 × 30	3.5 × 21	3.75 × 32	3.75 × 32	3.75 × 32
P	89 ± 1.5%	77.2 ± 2.9%	72.4 ± 3.1%	70.3 ± 2.8%	66.0 ± 2.6%
PD (pores/mm ²)	0.6 ± 0.5	6.3 ± 2.5	5.6 ± 1.4	8.2 ± 2.2	10.2 ± 2.9
Number of wires	16 struts	48 braided wires	Outer: 16 struts; inner: 36 braided wires	Outer: 16 struts; inner: 48 braided wires	Outer: 16 struts; inner: 64 braided wires

Note:—SSFD indicates stent-in-stent endoluminal flow diverting device.

level of the aneurysm neck. This phenomenon could perhaps explain some failures of flow diversion and the lack of neointima formation at the neck of treated aneurysms, with persisting aneurysm filling despite FD treatment.^{2,3}

In vitro studies that have hitherto been published have failed to uncover these variations because they studied stents and FDs in closed tubes.⁴ We here present the results of several simple experiments that permit a better understanding and more accurate prediction of the changes in structure of braided self-expanding stents and FDs as these adapt to the common types of geometric constraints encountered in treating aneurysms. This work can help inform the proper choice of stent and length of landing zone, and perhaps also illuminate some technical aspects of an intervention, in terms of the utility of the practice of intentional device compaction across the aneurysm neck.

MATERIALS AND METHODS

Materials

The main characteristics of studied stents and FDs are summarized in Table and illustrated in Fig 1. These included 5 devices: 1 HPS, 1 low-porosity FD (FD₄₈) similar to commercially available devices, and 3 prototype braided stent-in-stent endoluminal flow-diverting devices, composed of an outer HPS and an inner FD mesh of lower porosity and higher pore density, made with either 36, 48, or 64 wires, all gifts from MicroVention (Aliso Viejo, California). Experiments were performed with silicone tubes of various diameters (2, 2.5, and 3 mm), to simulate intracranial parent arteries and branches. Scaled microphotographs were taken for each experimental step with a stereo microscope and analyzed with image processing software (ImageJ; National Institutes of Health, Bethesda, Maryland).

Methods

In each experimental setup, the stent or FD was submitted to constraining conditions but also allowed to “react” to those conditions. The device extremities were 1) constrained within tubes of various diameters (simulating the parent artery or a branch of a bifurcation), and forced to follow geometries (curves of decreasing radii of curvature for example) or compacted by axial approx-




FIG 1. Devices. The first set of experiments is shown, in which various devices were used. Note the similar pattern shared by all braided devices, with a concentration of stent struts in the midportion of the device, compacted between relatively porous segments. The phenomenon, more marked with increasing compaction (as shown on the right side), leads to heterogeneities in porosity along the length of the same device.

imation of the tubes but 2) the device midsection was free to expand or adapt between constrained extremities (simulating the room made available at the level of a bifurcation and/or the neck of the aneurysm) (Fig 1). The pattern of strut distribution after each deformation was used to differentiate 3 different zones, which will be described. In each zone, parameters that were recorded included SF, defined as the number of struts per unit length, porosity (P), defined as the proportion of the surface area without metal coverage over total surface area, and pore density, the number of metal-enclosed pores per unit surface area. The effect of tube size was analyzed for each set of experiments. SF at different degrees of axial compression was determined by selecting a given stent segment and counting the number of struts per unit of length; and P was determined by measuring $(1 - \text{Total Metal Surface Area}) \times 100\%$. For a given device, there is a direct relationship between P, SF, and pore attenuation. P will be used as the variable for most reported results.




FIG 2. Zones. The 3 characteristic zones are illustrated in a severely compacted HPS (A). Progressive approximation of the tubes (axial compression) leads to decreasing porosities of the compaction Zone₃, but Zones₁ and ₂ remain relatively unchanged, as shown in B (HPS in 2-mm tubes). Axial compression leads directly to a shortening of Zone₃, but the transition Zone₂ length remains relatively unchanged (C). The same phenomenon was reproduced with the FD₄₈ (compare B and D), with porosities reaching 0 with severe compaction.

The experiments can be divided into 3 sets:

- 1) The first set was designed to test the impact of axial compression (0–7 mm) on strut distribution within the portion of the device free to react to such manipulations, the free segment of stent (or FSS). The experiments were repeated with extremities of the device inserted in tubes of 2, 2.5, and 3 mm in diameter. The hypothesis was that, because the surface of a cylinder can be estimated by the formula: $2\pi rl$ (where r and l are the radius and length of the cylinder), any decrease in the length available for metallic deployment, in the absence of a change in radius (changes do not occur in the midportion of the stent; see below), would automatically result in a proportional increase in SF or metallic attenuation and a decrease in porosity according to the formula: $SF_2/SF_1 = l_1/l_2 = 1 - P_2/1 - P_1$. This relationship was verified empirically.
- 2) The second set of experiments evaluated the effects of curvatures on porosity. Devices deployed into silicone tubes of various diameters at each extremity were bent to 0, 45, 90, and 135°. For each curvature, stents and FDs were also shortened bilaterally by 0–2 mm to measure the impact of axial compression combined with curvature on parameters

of interest. The stent strut distribution was compared between the convex and the concave surfaces of the FSS and at the level of various stent and FD zones. The hypothesis was that as the device follows the curve, the length available for the distribution of metal struts would differ between the convex and concave side of the curvature, with P and SF varying according to the ratio of the arc of the convex surface (C_v) over the arc of the concave surface (C_c): $SF_v/SF_c = C_v/C_c = 1 - P_c/1 - P_v$. The strength of the correlations between dependent variables (eg, distance between tubes, arc ratios) and resulting porosities were estimated by using the Pearson correlation coefficient R^2 .

- 3) The final set of experiments was designed to assess the behavior of the terminal ends of the device within the various sizes of silicone tubes, to simulate landing zones. The length of the landing zone was sequentially decreased, and the diameter of the terminal end of the device was measured. Various devices (HPS, FDs, stent-in-stent endoluminal flow-diverting devices with 48 or 64 wires) were deployed into silicone tubes of 2-, 2.5-, and 3-mm diameters. Experiments were repeated in glass tubes of the same diameter. Overall, 370 experiments were performed.

RESULTS

Zones

Placement of both ends of all devices into the tubes led to a redistribution of the struts according to 3 distinct zones, shown in Figs 1 and 2. The constrained or landing Zone₁ is the portion still inside the silicone tube; the unconstrained segment of the stent or FD, the FSS, can be further divided into 2 transition Zone₂, beginning immediately as the stent exits the tube, on both sides, and a middle expansion Zone₃, where the device has been allowed to expand completely, found between each transition Zone₂ (Fig 3, A–D). Both the inner and the outer stents of the stent-in-stent endoluminal flow-diverting devices displayed the same pattern (Fig 1, E–H).

Landing Zone₁

Constraining the device in tubes of progressively smaller diameter led to a small but linear decrease in porosity within the landing Zone₁ (for example, from 90% to $87 \pm 2\%$ for the HPS). Application of axial compression, bending, or combined axial compression with bending did not change the strut frequency or porosity of the device within this Zone₁ (Fig 2B, -D).

Transition Zone₂

The free segment of stent is composed of 2 transitions Zone₂ and 1 compaction Zone₃. The transition Zone₂ exists because of the difference in diameters between landing Zone₁ (constrained within the tube) and Zone₃ (free to expand). The transition Zone₂ is absent if the device has the same or smaller diameter than the tube (or parent vessel). With tubes of decreasing diameter, or with progressive relative oversizing of the device, the transition Zone₂ becomes longer (Pearson $R^2 = 0.99$) and an increasing proportion of the FSS is occupied by the more porous transition Zone₂ (Fig 3). In all cases, the porosity follows a gradient, progressive and continuous along Zone₂, between the maximum values found within the tube (Zone₁) and the minimal values found at Zone₃. Axial compression and progressive curving had minimal effect on the porosity of the transition Zone₂ (Figs 2–4).

Expansion Zone₃

The expansion Zone₃ is the only segment of the FSS that responds to compaction and angulation of the device. The 2 extremities of




FIG 3. Compaction and diameters. Effects of 2-mm (left column) and 4-mm compaction (right column) and of tube diameters (2–3.5 mm) on the compaction Zone₃ and the transition Zone₂ of the HPS are shown (A–H). Note that tubes of decreasing diameters lead to increased compaction of Zone₃ because the transition Zone₂ increases in length as tube diameter decreases. The relationship between the length of the transition Zone₂ and the size of tubes is illustrated in I for both the HPS and the FD₄₈.

Zone₃ consisted of 2 areas that appeared to be more dense in metallic coverage than the middle of Zone₃ (Figs 1 and 3). For simplicity, and because this distinction disappeared with increasing compaction, these will not be considered as individual entities. In contrast to the transition Zone₂, this portion of the stent underwent a significant decrease in porosity with approximation of the tubes. In keeping with our hypothesis, this fol-

lows a linear correlation with Pearson R^2 values of 0.93 or greater for all devices. With axial compression of 0–7 mm, the HPS porosity went from 80% to 72% and, for FDs, from 53% to 0% (Fig 2, B–D).

Bending the device from 0 to 135° increased porosity on the convex surface of the compaction Zone₃, while decreasing porosity on the concave surface (Fig 5). The change in porosity was directly proportional to the arc ratios, as hypothesized, with Pearson R^2 values of 0.95 for the HPS. Axial compression of the curved device, regardless of the degree of curvature, had the effect of “compacting” or increasing SF, thus decreasing the porosity of both the convex and the concave surfaces, but only within the compaction Zone₃; the transition Zone₂ and landing Zone₁ remained unchanged (Figs 4 and 5).

Minimal Length of the Landing Zone₁

With all FD devices deployed into silicone tubes, some degree of stenosis at the terminal ends could be observed. No such phenomenon was noted with the HPS or when we deployed devices into glass tubes. The stenosis increased from a minimal (4%) to severe (60%) stenosis as the length of the landing Zone₁ was progressively decreased, as shown in Fig 6. The stenosis-producing phenomenon became more pronounced with more marked mismatch between the diameter of the FDs and that of the constraining tubes. The stenosis at the ends of the device was more pronounced with devices of lesser nominal porosity. When a short landing zone was combined with a substantial angulation, the stenosis was asymmetrical (Fig 6).

DISCUSSION

The salient findings of this study are as follows:

- 1) Braided stents and FDs can exhibit major heterogeneities in porosity as these are subjected to various experimental conditions, including being simultaneously constrained into vessels of a smaller diameter than their nominal diameter but free to expand between the 2 constrained extremities.
- 2) The device adapts according to 3 characteristic zones.




FIG 4. Curvatures and FDs. Curvatures affect porosities of the convexity and concavity of FD₄₈. The effects are limited to the compaction Zone₃, the transition Zone₂ being relatively spared.





FIG 5. Concavity at 45°(A, C) and convexity at 135°(B, D) of expansion Zone₃ before (A, B), or after (C, D) the stent is compacted. Curvatures affect the porosity of the convexity and concavity sides of the compaction Zone₃ (E), whereas the transition Zone₂ remains virtually unchanged by comparison (F, showing the porosity of the transition Zone₂ with different tubes and different curvatures).

- 3) It is possible to compact stents and FDs and to a certain extent decrease porosities in hope of increasing efficacy of the stenting or FD strategy, but this works only at the level of the compaction Zone₃; a limiting factor is the transition Zone₂, of relatively higher porosity, which remains relatively unchanged.
- 4) Changes in porosity due either to deployment techniques




FIG 6. Stenoses can occur at the level of device extremities (A–D); the stenosis is more severe when the length introduced inside the tube is insufficient (compare A and B), and eccentric when the device is curved (D). The relationship between stenoses and length of device introduced inside tubes is shown for FD₄₈ (E) as well as for the 48-wire stent-in-stent endoluminal flow-diverting devices (F).

such as compaction or anatomic factors such as curvatures are predictable once it is recognized that changes are limited to the compaction Zone₃; these predictable alterations follow relatively simple geometric rules.

- 5) Stenoses at the level of the stent extremities can occur if the landing zone is too short.

We have tested high porosity braided stents and flow diverters in wide-neck, bifurcation, and giant aneurysm models, and failures have occurred, particularly when the device was deployed in a curved configuration.^{1–3} When in vivo porosities of the segments of devices bridging the aneurysm ostium were measured, it became clear that these were strongly influenced by discrepancies in diameters between the parent artery and the segment free to expand at the level of the aneurysm neck or bifurcation. Device deformations and subsequent heterogeneity in device porosity were frequent and sometimes severe.^{1–3} The invariably increased porosity at the transition Zone₂ may perhaps explain some failures of flow diversion, along with a lack of neointima formation across the neck of treated aneurysms and persistent filling after treatment.^{1,2,5}

The benchtop experiments presented here were designed to model in vivo findings we encountered after experimentation in a canine aneurysm model, as shown in Fig 7. Many of our observations were most obvious when there was a significant discrepancy between stent size and vessel lumen. Previous in vitro studies have emphasized the importance of matching the size of the device

with the diameter of the parent artery,⁴ and to this we are adding that device oversizing may also increase the length of the transition zone, a segment of device whose porosity is greater than the expansion zone. It is noteworthy that unlike the compaction zone, the porosity of the transition Zone₂ cannot be decreased with technical maneuvers such as axial compression.

The problems with device deployment into curved configurations have previously been studied, with an emphasis on changes in flow patterns, the angle progressively converting shear-driven into inertia-driven flows.^{6,7} The increased porosities on the convexity of an acute curve may also decrease the efficacy of a FD strategy.^{2,4} This has recently been shown in computational fluid dynamics models.⁷ Although the bench experiments can show the changing porosity of the device, the reality of deploying flow diverters within a curve is more complex. Transmitting force around a corner to get the device to open can cause problems when treating patients with FDs. Prior understanding of patient-specific blood vessel and aneurysm anatomy, along with required degree of curvature of the deployed endovascular device, would be important elements to consider when attempting to prescribe an amount of metallic coverage over an aneurysm or important branch ostium. Industry partners should develop charts or graphs of resultant device porosities, at each device diameter, depending on parent vessel diameter, and approximate curvature, to facilitate the prescriptive process.

Because the problems we describe here have now been repro-




FIG 7. In vivo and in vitro observations. Photographs of 4 autopsy specimens of aneurysms treated with HPSs (A, B) or FDs (C, D) are compared with 4 photographs of benchtop studies designed to mimic in vivo findings (E–H). Note how neointima formation (A–C) tends to be limited to the less porous compaction Zone₃, and how the transition Zone₂, lacking neointima, may be responsible for failures. The in vivo stent stenosis (D) is reproduced inside a tube (H).

duced with all devices we have tested, these seem to be intrinsic to their braided self-expandable nature. Ways to circumvent or palliate the problems can be imagined: intra-aneurysmal coils may prevent some expansion of the device at the level of the neck, or telescoping stents or flow diverters of progressively smaller diameters may decrease the discrepancies and minimize the phenomenon. These techniques should first be attempted in preclinical studies, because they may potentially increase other types of complications, such as parent vessel thrombosis.

Device stenosis occurring at the terminal ends of a device has been documented in some animal models, particularly when the device undergoes substantial expansion into the aneurysm neck (where a more marked discrepancy exists between diameters at the aneurysm ostium and the parent artery), with short landing zones (Fig 6). The present study suggests that this phenomenon may be prevented by choosing a long enough device to leave adequately long landing zones on either side of the aneurysm neck. Device manufacturers could also provide data on anticipated degree of device stenosis at each length of landing zone.

Limitations

Many devices used for this study were prototypes, and results may not apply to clinically available FDs. Recently, the HPS has been approved in Europe and Canada (LVIS; MicroVention). The FD₄₈ construction was modeled on the Pipeline device (ev3, Irvine, California). In vitro experiments were done in artificial environments, and although many phenomena have been documented in experimental animals, it is unclear whether and to what degree the same problems occur with currently approved intravascular devices in clinical applications. In vivo conditions may differ significantly from our experimental conditions, and any inference to human applications must be cautious.

CONCLUSIONS

Braided stents and FDs may exhibit heterogeneities in porosity that are to a certain extent predictable. Carefully matching device and parent vessel diameters and providing sufficient landing zones may prevent deformations that could affect the clinical results of stent placement and flow diversion in the treatment of complex aneurysms.

Disclosures: Jean Raymond—**RELATED:** Grant: FARQ/FRSQ, Comments: This work was partially supported by the Fondation de l'Association des Radiologistes du Québec (FARQ) in collaboration with Fonds de la Recherche en Santé du Québec (FRSQ) grant (to Dr. Jean Raymond). This work was also supported by a pilot project grant from the Society of Interventional Radiology (SIR) to Dr Tim Darsaut and by an imaging research bursary fellowship from Société Française de Radiologie to Dr Fabrice Bing; **Other:** Stents and FDs were gifts from MicroVention.

REFERENCES

1. Darsaut TE, Bing F, Gevry G, et al. Flow diverters can occlude aneurysms and preserve arterial branches: a new experimental model. *AJNR Am J Neuroradiol* 2012;33:2004–09
2. Darsaut TE, Bing F, Salazkin I, et al. Flow diverters fail to occlude experimental bifurcation or curved side-wall aneurysms: an in vivo study in canines. *J Neurosurg* 2012;117:37–44
3. Darsaut TE, Bing F, Salazkin I, et al. Testing flow diverters in giant fusiform aneurysms: a new experimental model can show leaks responsible for failures. *AJNR Am J Neuroradiol* 2011;32:2175–79
4. Aurboonyawat T, Blanc R, Schmidt P, et al. An in vitro study of Silk stent morphology. *Neuroradiology* 2011;53:659–67
5. Byrne JV, Beltechi R, Yarnold JA, et al. Early experience in the treatment of intra-cranial aneurysms by endovascular flow diversion: a multicentre prospective study. *PLoS One* 2010;5:e12492
6. Meng H, Wang Z, Kim M, et al. Saccular aneurysms on straight and curved vessels are subject to different hemodynamics: implications of intravascular stenting. *AJNR Am J Neuroradiol* 2006;27:1861–65
7. Mut F, Cebal JR. Effects of flow-diverting device oversizing on hemodynamics alteration in cerebral aneurysms. *AJNR Am J Neuroradiol* 2012;33:2010–16