

Providing Choice & Value

Generic CT and MRI Contrast Agents





Differentiation of Hemorrhage from Iodine Using Spectral Detector CT: A Phantom Study

S. Van Hedent, N. Große Hokamp, K.R. Laukamp, N. Buls, R. Kessner, B. Rose, P. Ros and D. Jordan

This information is current as of July 15, 2025.

AJNR Am J Neuroradiol 2018, 39 (12) 2205-2210 doi: https://doi.org/10.3174/ajnr.A5872 http://www.ajnr.org/content/39/12/2205

Differentiation of Hemorrhage from Iodine Using Spectral Detector CT: A Phantom Study

^{(D}S. Van Hedent, ^{(D}N. Große Hokamp, ^{(D}K.R. Laukamp, ^{(D}N. Buls, ^{(D}R. Kessner, ^{(D}B. Rose, ^{(D}P. Ros, and ^{(D}D. Jordan

ABSTRACT

BACKGROUND AND PURPOSE: Conventional CT often cannot distinguish hemorrhage from iodine extravasation following reperfusion therapy for acute ischemic stroke. We investigated the potential of spectral detector CT in differentiating these lesions.

MATERIALS AND METHODS: Centrifuged blood with increasing hematocrit (5%–85%) was used to model hemorrhage. Pure blood, blood-iodine mixtures (75/25, 50/50, and 25/75 ratios), and iodine solutions (0–14 mg I/mL) were scanned in a phantom with attenuation ranging from 12 to 75 HU on conventional imaging. Conventional and virtual noncontrast attenuation was compared and investigated for correlation with calculation of relative virtual noncontrast attenuation. Values for all investigated categories were compared using the Mann-Whitney *U* test. Sensitivity and specificity of virtual noncontrast, relative virtual noncontrast, conventional CT attenuation, and iodine quantification for hemorrhage detection were determined with receiver operating characteristic analysis.

RESULTS: Conventional image attenuation was not significantly different among all samples containing blood (P > .05), while virtual noncontrast attenuation showed a significant decrease with a decreasing blood component (P < .01) in all blood-iodine mixtures. Relative virtual noncontrast values were significantly different among all investigated categories (P < .01), with correct hemorrhagic component size estimation for all categories within a 95% confidence interval. Areas under the curve for hemorrhage detection were 0.97, 0.87, 0.29, and 0.16 for virtual noncontrast, relative virtual noncontrast, conventional CT attenuation, and iodine quantification, respectively. A \geq 10-HU virtual noncontrast, \geq 20-HU virtual noncontrast, \geq 40% relative virtual noncontrast, and combined \geq 10-HU virtual noncontrast attenuation threshold had a sensitivity/specificity for detecting hemorrhage of 100%/23%, 89%/ 95%, 100%/82%, and 100%/100%, respectively.

CONCLUSIONS: Spectral detector CT can accurately differentiate blood from iodinated contrast in a phantom setting.

ABBREVIATIONS: ICH = intracranial hemorrhage; R-VNC = relative virtual noncontrast; SDCT = spectral detector CT; VNC = virtual noncontrast

Material with similar attenuation can be difficult to distinguish on conventional CT. A common clinical illustration of this problem is the differentiation of iodine from hemorrhage because both are hyperdense on conventional unenhanced CT. In patients with acute ischemic stroke, intra-arterial thrombolytic therapy has been shown to decrease morbidity and mortality.¹ However, it has also been reported that this increases the risk of intracranial hemorrhage (ICH), with reported frequencies of 10%–15% and mortality up to 83% for symptomatic ICH.²⁻⁷ Because of the disruption of the blood-brain barrier, contrast extravasation can occur during the procedure in 30%–50% of cases, impairing the detection of or the differentiation from ICH due to the overlap in density.^{5,8-11} As of this writing, unenhanced conventional CT is performed within 24 hours after treatment for the detection of complications. In case of unclear findings (eg, if differentiation between ICH and iodine extravasation is not possible), follow-up imaging may be performed. Hence, the ability to differentiate these 2 is highly desirable to avoid additional examinations and to ensure appropriate management.²

Please address correspondence to Steven Van Hedent, Universitair Ziekenhuis Brussel, Department of Radiology, Laarbeeklaan 101, 1090 Brussels, Belgium; e-mail: steven.vanhedent@case.edu

Received June 3, 2018; accepted after revision August 28.

From the Departments of Radiology (S.V.H., N.G.H., K.R.L., R.K., P.R., D.J.) and Pathology (B.R.), University Hospitals Cleveland Medical Center, Cleveland, Ohio; Case Western Reserve University School of Medicine (S.V.H., N.G.H., K.R.L., R.K., B.R., P.R., D.J.), Cleveland, Ohio; Institute for Diagnostic and Interventional Radiology (N.G.H., K.R.L.), University Hospital Cologne, Cologne, Germany; Vrije Universiteit Brussel (S.V.H., N.B.), Brussels, Belgium; and Department of Radiology (S.V.H., N.B.), Universitair Ziekenhuis Brussel, Brussels, Belgium.

Preliminary results previously presented as an electronic poster at: Annual Meeting of the European Congress of Radiology, February 28 to March 4, 2018;Vienna, Austria.

This work was partially supported by Philips Healthcare under a research agreement with University Hospitals Cleveland Medical Center and Case Western Reserve University.

http://dx.doi.org/10.3174/ajnr.A5872



FIG 1. Comparison of the attenuation of diluted blood, iodine, and blood-iodine mixtures on conventional and virtual noncontrast images. There is an incremental decrease of VNC attenuation values with decreasing blood content, compared with conventional CT attenuation values.

Two major discriminators of attenuation within a voxel are the energy of the x-ray beam and the concentration of attenuating material in that voxel. The presence or concentration of each discriminator in that voxel therefore cannot be determined by performing a single attenuation measurement from a broad photon energy spectrum. However, in dual-energy CT, attenuation at different energies is registered. In single- and dual-source dual-energy Revolution CT, this is achieved using a fast-kilovolt(peak) switching x-ray source (GE Healthcare, Milwaukee, Wisconsin) or 2 x-ray sources and 2 detectors (Somatom Force, Siemens Healthineers, Forchheim, Germany), respectively. Spectral detector CT (SDCT) (IQon; Philips Healthcare, Best, the Netherlands) distinguishes low- and high-energy data at the level of the detector using a dual-layer detector. The bottom and top layer absorb high- and low-energy photons, respectively. The main advantage of the latter approach is that spectral data are collected without the need to prospectively choose a spectral scanning protocol or mode, as needed in other dual-energy CT approaches.

Previous studies have shown the potential of dual-energy CT for differentiating iodine from hemorrhage in clinical settings^{10,12-16}; however, no study has been performed using SDCT for this application nor has the sensitivity or specificity been determined in a phantom system.

In our study, we investigated the ability of SDCT to differentiate ICH from iodinated contrast in a phantom model.

MATERIALS AND METHODS

Image Acquisition

All scans were performed with a 464 Phantom (Gammex, Madison, Wisconsin) with a plastic 50-mL tube in its air cavity, filled with samples simulating hemorrhage (blood), diluted iodine, and a hemorrhage-iodine (blood-iodine) mixture. Each sample was scanned in the phantom with the following settings on the IQon SDCT scanner: 200 mAs, 120 kVp, pitch = 1, gantry rotation time = 330 ms, detector collimation = 64×0.625 mm, volume CT dose index = 18 mGy. Conventional CT images were reconstructed using an iterative reconstruction algorithm (iDose 4, Level 3; Philips Healthcare), while the spectral images were reconstructed using a spectral reconstruction algorithm (Spectral B, level 0). Reconstructed slice thickness was 3 mm. All scans were performed 3 times to account for interscan variability.

Sample Preparation

Previous studies have shown that the density of blood is primarily determined by the hemoglobin concentration.¹⁷ In turn, changes in hematocrit are reflected in the density of a hemorrhage seen on conventional CT.18 We collected discarded packed red blood cells and prepared samples to a range of 0%, 5%, 15%, 25%, 35%, 45%, 55%, 65%, 75%, and 85% hematocrit. The packed red blood cells were prepared using phosphate buffered saline to prevent cytolysis. Each sample was scanned with the protocol described above. Diluted iodine (Optiray 350, Ioversol; Mallinckrodt, St. Louis, Missouri) samples in

phosphate buffered saline, with concentrations ranging from 0 to 14 mg I/mL (0.00, 0.70, 1.40, 2.10, 2.80, 3.50, 4.20, 4.90, 5.60, 7.00, 14.00 mg I/mL) were scanned using the same protocol. We matched the iodine attenuation with the densities found for 85%, 75%, 65%, 55%, 45%, and 35% hematocrit to create iodine dilutions with matching densities (2.50, 2.10, 1.80, 1.40, 1.10, and 0.70 mg I/mL, respectively) at those hematocrit levels.

Blood-iodine mixtures with matching densities were scanned with following proportions: [2/3] blood + [1/3] iodine; + [1/2] blood + [1/2] iodine; [1/3] blood + [2/3] iodine.

Image Analysis

All image analysis was performed using the proprietary image viewer (Spectral Diagnostic Suite; Philips Healthcare) of the vendor. Each diluted blood, diluted iodine, and blood-iodine mixture sample was analyzed by placing a circular 2-cm² ROI centrally in the inserts. The attenuation on the conventional and virtual noncontrast (VNC) image (using VNC image) and iodine concentration (using the iodine density map) were measured within each ROI (Fig 1). Relative VNC (R-VNC) attenuation (%) is calculated using the following equation:

$$\text{R-VNC} = \frac{\text{Attenuation VNC}}{\text{Attenuation Conventional}} \times 100.$$

Statistical Analysis

Statistical analysis was performed using SPSS 21.0 (IBM, Armonk, New York). Iodine-quantification measurements were compared with true iodine concentrations for correlation. Mean iodine quantification error (milligram/milliliter) was calculated and presented with a 95% confidence interval and Bland-Altman plot analysis. Attenuation in the conventional and VNC images was compared for correlation using the Pearson correlation. Attenuation on conventional images, VNC attenuation, and R-VNC are reported with 95% CIs for the investigated sample compositions (blood, [2/3] blood + [1/3] iodine; [1/2] blood + [1/2] iodine; [1/3] blood + [2/3] iodine, iodine) and compared by means of the Mann-Whitney *U* test. VNC attenuation for the samples containing blood was compared with the hematocrit in the samples for correlation. VNC attenuation, R-VNC attenuation, and attenuation on the conventional images were analyzed for sensitivity and specificity in detecting a hemorrhagic component. Receiver operating characteristic analysis was performed for conventional CT, VNC, R-VNC attenuation, and iodine quantification, and the corresponding areas under the curve with 95% CIs are reported. The difference among areas under the curve was assessed by the method of Hanley and McNeil.¹⁹ A *P* value < .05 was considered statistically significant.

RESULTS

Mean attenuation on the conventional and VNC images (\pm 95% CI) for the pure blood samples was similar: respectively, 39.1 \pm 31.5–46.7 HU and 39.1 \pm 31.5–46.7 HU (Fig 2 and the Table). The blood-iodine mixtures, consisting of [2/3] blood + [1/3] iodine; [1/2] blood + [1/2] iodine; [1/3] blood + [2/3] iodine had almost identical attenuation values on conventional images: 49.48 \pm 42.8–56.2 HU, 49.2 \pm 41.7–56.7 HU, and 49.4 \pm 42.2–56.6 HU, respectively. Meanwhile, the corresponding VNC attenuation values incrementally decreased with an increasing iodine fraction, measuring 40.4 \pm 35.9–44.9 HU, 32.6 \pm 28.6–36.6 HU, and 26.0 \pm 23.3–28.6 HU, respectively.

Differences in attenuation on the conventional images among the different investigated compositions containing a blood component were not significant (P > .05). Conversely, VNC attenuation values were significantly different among samples containing 67% blood, 50% blood, and 33% blood, but not among samples consisting of 100% and 67% blood (P = .70).

Conventional and VNC attenuation correlated significantly



FIG 2. Attenuation values (HU) of diluted blood, blood-iodine mixtures, and diluted iodine on spectral detector CT conventional and virtual noncontrast images. There is a significant incremental decrease of VNC attenuation values with decreasing blood content. The *asterisk* indicates a significant difference of conventional CT attenuation compared with other compositions (P < .0); double asterisks, significant differences of VNC attenuation among these compositions (P < .0).

for all samples containing a blood component ($R^2 > 0.97$, P < .01). Correlation was highest for pure blood dilutions ($R^2 = 1.00$), followed by [2/3] blood + [1/3] iodine, [1/2] blood + [1/2] iodine mixtures, and it was lowest for samples containing [1/3] blood + [2/3] iodine ($R^2 = 0.97$).

Correlation between the hematocrit level in our samples and VNC attenuation was excellent ($R^2 = 0.97, P < .01$), while correlation with the attenuation on the conventional images was moderate, but still significant ($R^2 = 0.50, P < .01$) (Fig 3).

For the samples consisting only of diluted iodine, mean attenuation on conventional and VNC images (\pm 95% CI) was 129.5 \pm 99.4–159.6 HU and 14.5 \pm 13.4–15.6 HU, respectively, which were both significantly different from other samples containing a blood component (P < .01) (Fig 2 and the Table). When one interprets these results, it is important to note that phosphate buffered saline has different attenuation properties from normal saline. The average attenuation of samples filled with only phosphate buffered saline was 8.5 and 8.3 HU on conventional and VNC images, respectively. Correlation between conventional and VNC attenuation for samples with pure iodine dilutions was not significant ($R^2 = 0.77$, P = .08).

The differences between R-VNC attenuation values of all compositions were significant (P < .01), as shown in Fig 4 and the Table. Mean R-VNC attenuation (\pm 95% CI) for the pure blood, [2/3] blood - [1/3] iodine, [1/2] blood - [1/2] iodine, [1/3] blood - [2/3] iodine, and pure iodine was 99.59% \pm 98.8%-100.4%, 82.8% \pm 79.5%-86.1%, 67.6% \pm 64.4%-70.9%, 54.3% \pm 50.6%-58.0%, and 15.7% \pm 12.4%-19.1%. There was no overlap within the 95% CI between R-VNC values for all compositions, contrary to VNC values in which all samples containing blood showed overlap.

There was a significant correlation between measured and true iodine concentrations ($R^2 > 0.99$, P < .01), with a mean iodine quantification error of $-0.41 \pm 0.31-0.50$ mg/mL) (Fig 5).

Receiver operating characteristic curve analysis for hemorrhagic-component detection showed a significant difference among areas under the curve (P < .01), being highest for VNC attenuation ($0.97 \pm 0.94-0.99$ HU), followed by R-VNC attenuation ($0.87 \pm 0.77-0.97$ HU) and attenuation in the conventional CT images ($0.29 \pm 0.16-0.41$ HU), and lowest for iodine quantification ($0.16 \pm 0.06-0.25$ HU) (Fig 6*A*); these differences were significant (P < .01). Using a threshold of ≥ 10 and ≥ 20 HU for VNC had a sensitivity/specificity of 100%/23% and 89%/95%, respectively. An R-VNC attenuation of $\geq 40\%$ had a sensitivity/ specificity of 100%/82%. Conversely, using a threshold of ≥ 10 and ≥ 20 HU on the conventional CT images had a sensitivity/ specificity of 100%/13% and 94%/15%, respectively (Fig 6).

Attenuation values (HU) of diluted blood	blood-iodine mixtures, and diluted i	iodine on SDCT conventiona	l and VNC images ^a
--	--------------------------------------	----------------------------	-------------------------------

	Conventional (HU)		VNC (HU)		R-VNC (%)	
	Mean	95% CI	Mean	95% CI	Mean	95% CI
100% blood–0% iodine	39.1	31.5-46.7	39.1	31.5-46.7	99.6	98.8–100.4
67% blood–33% iodine	49.5	42.8-56.2	40.4	35.9-44.9	82.8	79.5–86.1
50% blood–50% iodine	49.2	41.7-56.7	32.6	28.6-36.6	67.6	64.4–70.9
33% blood–67% iodine	49.4	42.2–56.6	26.0	23.3-28.6	54.3	50.6-58.0
0% blood –100% iodine	129.5	99.4–159.6	14.5	13.4–15.6	15.7	12.4–19.1

^a R-VNC attenuation (%) was calculated by comparing VNC attenuation with the attenuation on conventional CT for all investigated categories.



FIG 3. Correlation between the hematocrit in our dilutions and the attenuation in the conventional and VNC images.



FIG 4. Relative VNC attenuation (%), by comparing VNC attenuation with the attenuation on conventional CT for all investigated categories (diluted blood, blood-iodine mixtures, and diluted iodine). R-VNC values among all compositions were significantly different (P < .01).

DISCUSSION

Our results show that SDCT has excellent diagnostic accuracy for differentiating blood from iodinated contrast in a phantom. We observed similar sensitivity and specificity (both >90%) for hemorrhage detection using spectral dual-layer VNC images, compared with previous clinical studies that used single- or dualsource dual-energy CT.^{10,13,15} These studies used visual assessment by radiologists for classifying hyperdensities on "simulated" conventional head CT images as hemorrhage or iodine; hyperdensities visible on VNC images were classified as having a hemorrhagic component. Although typical ICHs are hyperdense (>50 HU), they can be associated with a lower density due to anticoagulation, the presence of CSF with arachnoid laceration, or severe anemia (eg, sickle cell anemia), which can complicate

Virtual noncontrast

subjective assessment.²⁰⁻²⁹ Our results show high accuracy by quantitative VNC assessment for blood detection, including low densities on conventional images (as low as 10 HU). By combining a \geq 10 HU VNC and \geq 40% R-VNC cutoff, we observed a sensitivity and specificity of 100% for blood, including low densities and mixtures with iodine on conventional images (Fig 6).

Several authors have investigated the accuracy of SDCT for material decomposition, which allows iodine to be subtracted from an image. These studies have shown high accuracy for iodinequantification and VNC attenuation values, though results vary. Our results show slightly lower iodine-quantification accuracy than recent publications, which can be explained by the extremely low iodine concentrations needed in our study, though results are still excellent. We had results comparable with those of Pelgrim et al,³⁰ which showed a median

error of -0.6 mg/mL, while more recent studies showed differences ranging from -0.46 to 0.1 mg/mL.³¹ Regarding VNC, Duan et al³¹ showed good agreement between measured spectral detector VNC attenuation values and reference standards (-9.95-6.41 HU), confirmed by our study. Still, VNC inaccuracies can occur. When we incorporated a comparison of the attenuation between VNC and conventional images, small inaccuracies of true VNC values can be negated, which can explain the excellent results of R-VNC in our study for hemorrhage detection (100% sensitivity and 82% specificity) and size estimation (Figs 4 and 5).

Clinical studies using SDCT differentiating iodine from blood in skull imaging have been rare. As of this writing, the authors found only a pilot study from Cho et al,32 showing that spectral data analysis can be helpful in discriminating ICH from contrast enhancement in intracranial malignancies. Regarding the detection of hemorrhage, a recent study of Nute et al³³ used singlesource dual-energy CT for distinguishing ICH from calcification in a phantom model with densities from 40 to 100 HU, resulting in an accuracy of >90%.

There are several limitations to our study. First, our results are without clinical data because this was not within the scope of our investigation. Our goal was to investigate in a phantom setting. Second, our hemorrhage dilutions were prepared with phosphate buffered saline to prevent red blood cell hemolysis, but this could potentially bias the attenuation characteristics of blood and iodine in our results: Phosphate buffered saline is not representative of brain tissue in a clinical setting. Third, our mixtures were homogeneously mixed for optimal spectral analysis, which may not always be the case in a clinical setting. Last, we used a relatively large 2-cm² ROI for our measurements, so our results could not be affected by spatial resolution and partial volume limitations. This might not always be feasible in patients if the size of the clinical ROI is small.



FIG 5. Results of the iodine quantification measurements by comparison of measured-to-true iodine concentrations (A) and errors in iodine quantification measurements by a Bland-Altman plot (B). A, Correlation between measured and true iodine concentrations is excellent ($R^2 > 0.99$, P < .01). B, Mean iodine quantification error (\pm 95% CI) was $-0.41 \pm 0.31-0.50$ mg/mL).



FIG 6. Performance of conventional CT attenuation, virtual noncontrast attenuation, and relative VNC attenuation for the detection of blood. *A*, Receiver operating characteristic curve analysis shows the highest area under the curve for VNC ($0.97 \pm 0.94-0.99$), followed by R-VNC attenuation ($0.87 \pm 0.77-0.97$) and attenuation in the conventional CT images ($0.29 \pm 0.16-0.41$), and the area under the curve was lowest for iodine quantification ($0.16 \pm 0.06-0.25$). *B*, When we combined a $\geq 40\%$ R-VNC (*dashed line*) and ≥ 10 HU VNC cutoff, there is 100% differentiation between blood-containing (diluted blood and blood-iodine mixtures) and diluted iodine samples.

CONCLUSIONS

Our results show that the added spectral information of SDCT has high sensitivity and specificity in detecting blood and can accurately estimate hemorrhagic component size, including when mixed with iodine. This information could be of potential benefit in brain imaging for patients following reperfusion therapy in acute stroke.

Disclosures: Steven Van Hedent—RELATED: Grant: Philips Healthcare, Comments: Partial funding support was provided by Philips Healthcare under a research agreement with University Hospitals Cleveland Medical Center and Case Western Reserve University.* Nils Grosse Hokamp—RELATED: Grant: University Hospitals Cleveland Medical Center/Case Western Reserve University/Philips Healthcare, Comments: Part of this Research was funded under a research agreement among University Hospitals Cleveland Medical Center, Case Western Reserve University, and Philips Healthcare*; UNRELATED: Payment for Lectures Including Service on Speakers Bureaus: Philips Healthcare, Comments: Speakers Bureau; Payment for Development of Educational Presentations: Philips Healthcare. Nico Buls—UNRELATED: Payment for Lectures Including Service on Speakers Bureaus: GE Healthcare, Comments: lectures on dual-energy CT.* Pablo Ros—UNRELATED: Other: Philips Healthcare, Comments: The CT unit where this research was performed is provided by Philips Healthcare to our hospital to test clinical applications. This project was an exhibit in our existing Master Research Agreement. No direct funds were received*; *OTHER RELATIONSHIPS*: hospital grant from Philips regulated by Master Research Agreement, consisting of a Spectral Detector CT unit. David Jordan—*RELATED*: *Grant*: Philips Healthcare, *Comments*: The IQon CT scanner was provided to University Hospitals Cleveland Medical Center under a research agreement*; *UNRELATED*: *Payment for Lectures Including Service on Speakers Bureaus*: Medical Technology Management Institute/Herzing University, *Comments*: paid speaker for continuing education programs for medical physicists. Kai Roman Laukamp—*RELATED*: *Grant*: Philips Healthcare, *Comments*: Philips Healthcare provided partial funding*. *Money paid to the institution.

REFERENCES

- National Institute of Neurological Disorders and Stroke rt-PA Stroke Study Group. Tissue plasminogen activator for acute ischemic stroke. N Engl J Med 1995;333:1581–87 CrossRef Medline
- Mokin M, Kan P, Kass-Hout T, et al. Intracerebral hemorrhage secondary to intravenous and endovascular intraarterial revascularization therapies in acute ischemic stroke: an update on risk factors, predictors, and management. *Neurosurg Focus* 2012;32:E2 CrossRef Medline
- 3. Fiorelli M, Bastianello S, von Kummer R, et al. Hemorrhagic transformation within 36 hours of a cerebral infarct: relationships with early clinical deterioration and 3-month outcome in the European

Cooperative Acute Stroke Study I (ECASS I) cohort. *Stroke* 1999;30: 2280–84 CrossRef Medline

- Nakano S, Iseda T, Kawano H, et al. Parenchymal hyperdensity on computed tomography after intra-arterial reperfusion therapy for acute middle cerebral artery occlusion: incidence and clinical significance. *Stroke*2001;32:2042–48 CrossRef Medline
- Yoon W, Seo JJ, Kim JK, et al. Contrast enhancement and contrast extravasation on computed tomography after intra-arterial thrombolysis in patients with acute ischemic stroke. *Stroke* 2004;35: 876–81 CrossRef Medline
- 6. Yedavalli V, Sammet S. Contrast extravasation versus hemorrhage after thrombectomy in patients with acute stroke. *J Neuroimaging* 2017;27:570–76 CrossRef Medline
- Greer DM, Koroshetz WJ, Cullen S, et al. Magnetic resonance imaging improves detection of intracerebral hemorrhage over computed tomography after intra-arterial thrombolysis. *Stroke* 2004; 35:491–95 CrossRef Medline
- Jang YM, Lee DH, Kim HS, et al. The fate of high-density lesions on the non-contrast CT obtained immediately after intra-arterial thrombolysis in ischemic stroke patients. *Korean J Radiol* 2006;7: 221–28 CrossRef Medline
- Mericle RA, Lopes DK, Fronckowiak MD, et al. A grading scale to predict outcomes after intra-arterial thrombolysis for stroke complicated by contrast extravasation. *Neurosurgery* 2000;46:1307–14; discussion 14–15 CrossRef Medline
- Gupta R, Phan CM, Leidecker C, et al. Evaluation of dual-energy CT for differentiating intracerebral hemorrhage from iodinated contrast material staining. *Radiology* 2010;257:205–11 CrossRef Medline
- Yokogami K, Nakano S, Ohta H, et al. Prediction of hemorrhagic complications after thrombolytic therapy for middle cerebral artery occlusion: value of pre- and post-therapeutic computed tomographic findings and angiographic occlusive site. *Neurosurgery* 1996;39:1102–07 CrossRef Medline
- 12. Siddiqui FM, Bekker SV, Qureshi AI. Neuroimaging of hemorrhage and vascular defects. *Neurotherapeutics* 2011;8:28–38 CrossRef Medline
- 13. Tijssen MP, Hofman PA, Stadler AA, et al. The role of dual energy CT in differentiating between brain haemorrhage and contrast medium after mechanical revascularisation in acute ischaemic stroke. *Eur Radiol* 2014;24:834–40 CrossRef Medline
- Nute JL, Le Roux L, Chandler AG, et al. Differentiation of low-attenuation intracranial hemorrhage and calcification using dual-energy computed tomography in a phantom system. *Invest Radiol* 2015;50:9–16 CrossRef Medline
- Phan CM, Yoo AJ, Hirsch JA, et al. Differentiation of hemorrhage from iodinated contrast in different intracranial compartments using dual-energy head CT. AJNR Am J Neuroradiol 2012;33:1088–94 CrossRef Medline
- Hu R, Padole A, Gupta R. Dual-energy computed tomographic applications for differentiation of intracranial hemorrhage, calcium, and iodine. *Neuroimaging Clin N Am* 2017;27:401–09 CrossRef Medline
- 17. New PF, Aronow S. Attenuation measurements of whole blood and

blood fractions in computed tomography. *Radiology* 1976;121(3 Pt. 1):635–40 CrossRef Medline

- Bergström M, Ericson K, Levander B, et al. Variation with time of the attenuation values of intracranial hematomas. J Comput Assist Tomogr 1977;1:57–63 CrossRef Medline
- Hanley JA, McNeil BJ. The meaning and use of the area under a receiver operating characteristic (ROC) curve. *Radiology* 1982;143: 29–36 CrossRef Medline
- 20. Grelat M, Madkouri R, Bousquet O. Acute isodense subdural hematoma on computed tomography scan–diagnostic and therapeutic trap: a case report. J Med Case Rep 2016;10:43 CrossRef Medline
- Lee KS, Bae WK, Bae HG, et al. The computed tomographic attenuation and the age of subdural hematomas. *J Korean Med Sci* 1997;12: 353–59 CrossRef Medline
- Boyko OB, Cooper DF, Grossman CB. Contrast-enhanced CT of acute isodense subdural hematoma. *AJNR Am J Neuroradiol* 1991; 12:341–43 Medline
- Mendonça R, Lima TT, Dini LI, et al. Bilateral isodense epidural hematoma: case report. Arq Neuropsiquiatr 2005;63:862–63 CrossRef Medline
- May PL, Miles JB. Acute isodense extradural haematoma. Br J Neurosurg 1989;3:221–24 CrossRef Medline
- Rieth KG, Schwartz FT, Davis DO. Acute isodense epidural hematoma on computed tomography. J Comput Assist Tomogr 1979;3: 691–93 CrossRef Medline
- 26. Harbury OL, Provenzale JM, Barboriak DP. Vertex epidural hematomas: imaging findings and diagnostic pitfalls. Eur J Radiol 2000;36:150–57 CrossRef Medline
- 27. Claes F, Verhagen CV, Verhagen WI, et al. Acute isodense intracerebral haematoma due to coagulopathy associated with prostate cancer. *Clin Neurol Neurosurg* 2007;109:520–22 CrossRef Medline
- Kaufman HH, Singer JM, Sadhu VK, et al. Isodense acute subdural hematoma. J Comput Assist Tomogr 1980;4:557–59 CrossRef Medline
- 29. Thust SC, Burke C, Siddiqui A. **Neuroimaging findings in sickle cell disease.** *Br J Radiol* 2014;87:20130699 CrossRef Medline
- Pelgrim GJ, van Hamersvelt RW, Willemink MJ, et al. Accuracy of iodine quantification using dual energy CT in latest generation dual source and dual layer CT. Eur Radiol 2017;27:3904–12 CrossRef Medline
- 31. Duan X, Arbique G, Guild J, et al. Technical note: quantitative accuracy evaluation for spectral images from a detector-based spectral CT scanner using an iodine phantom. *Med Phys* 2018;45:2048–53 CrossRef Medline
- 32. Cho SB, Baek HJ, Ryu KH, et al. Initial clinical experience with duallayer detector spectral CT in patients with acute intracerebral haemorrhage: a single-centre pilot study. *PLoS One* 2017;12: e0186024 CrossRef Medline
- 33. Nute JL, Jacobsen MC, Chandler A, et al. Dual-energy computed tomography for the characterization of intracranial hemorrhage and calcification: a systematic approach in a phantom system. *Invest Radiol* 2017;52:30–41 CrossRef Medline